



NOTE:
REFER TO DATA SHEET TO SEE WHICH OPTIONS ARE PRESENT.

E	ADDED PURGE PORT; IDENTIFIED PURGE PORTS; UPDATED FOR NEW ENCLOSURE CONFIGURATION	GAR	5-18-06	THERMCO INSTRUMENT CORPORATION	SCALE: N.T.S. B. M. NUMBER:	DRAWN BY: DB
D	OPTIONAL PUMP CHANGED	GAR	5-27-05		DATE: 8-31-93 CUSTOMER:	APPROVED BY:
C	ADDED OPTIONAL FLOWMETER HOLES	DMR	12-30-04	S. O. NUMBER:	SUPERSEDES: 2-10133	
B	REPLACED PWR SUP 7805 WITH NEW ASSY INCLUDING NEW FUSE INFO (FUSE WAS 0.5 AMP); 115 WAS 110; ADDED 100 TO 115 VAC TRANSFORMER INFO; ADDED PURGE FLOWMETER INFO TAGS; ADDED MOUNTING STUD INFO; ADDED LABELS FOR POWER SWITCH AND ZERO POT; ADDED "SPAN POT"; ADDED SAMPLE GAS INLET INFO	GAR	5-21-02	TITLE: INSTRUCTION DRAWING FOR MODEL 7000 ANALYZER	DRAWING NUMBER: 2-10686	
A	ADDED OPTIONAL STEP-DOWN TRANSFORMER INFORMATION; CHANGED TO KEY FLOWMETERS; CORRECTED HEATER CONTROL MODULE	GAR	6-10-99			
	REVISIONS	BY	DATE			